



Model:

B-383LD2

Typology:

ROUTINE MICROSCOPE

Description:

Educational and laboratory microscope for routine applications.

Dye-cast frame, with high stability and ergonomy, for transmitted light and fluorescence reflected light observation.

Illumination	<p><u>Transmitted light:</u> Light source type X-LED³ with manual brightness control.</p> <p><u>Epi-fluorescence:</u> High-power white LED.</p>
Observation modes	<p>Brightfield, Fluorescence Fluorescence B: EX 460-490, DM 505, EM 515LP Fluorescence G: EX 510-550, DM 570, EM 590LP</p>
Filter Set	<p>2 positions fluorescence filter holder:</p> <p>Excitation B: Acridine Yellow, Acridine Orange, Auramine, DiO, DTAF, FITC, GFP, YFP, ecc. Excitation G: DiI; Blu Evans, Feulgen, Rhodamine, Texas Red, TRITC, PI, etc.</p>
Focusing	<p>Coaxial coarse and fine focusing mechanism (graduated, 0.002mm) with upper stop, to prevent the contact between objective and specimen. Adjustable tension of coarse focusing knob.</p>
Stage	<p>Double layer mechanical sliding stage, dimensions 216x150mm, moving range 78x54mm. Belt-drive in X direction. Vernier scale on the two axes, accuracy 0,1 mm.</p>
Nosepiece	<p>Quintuple revolving nosepiece, rotation on ball bearings.</p>
Head	<p>Trinocular observation head, inclined 30° and rotatable 360°. Diopter adjustment on both eyepieces. Interpupillary adjustment 48-75 mm. Fixed photo port, 50/50 ratio.</p>
Eyepieces	<p>Widefield eyepieces WF10X/20 with field number 20.</p>
Objectives	<p>E-PLAN objectives IOS (infinity corrected), made by: -) IOS 4X, A.N. 0.10, W.D. 16,8 mm -) IOS 10X, A.N. 0.25, W.D. 5,8 mm -) IOS 20x, A.N., 0,40, W.D. 1,4 mm -) IOS 40X, A.N. 0.65, W.D. 0,43 mm -) IOS 50X, A.N. 0,75, W.D. 0,3 mm (no cover slide) All objectives are treated with an anti-fungus treatment.</p>
Condenser	<p>Abbe condenser, N.A. 1,25 with centering system.</p>
Dimensions	<p>HEIGHT: 480 mm (without attachment) / 592 mm (with attachment) WIDTH: 210 mm DEPTH: 370 mm WEIGHT: 5 kg</p>
Accessories	<p>Instruction manual and dust cover included.</p>